




# TEST GAS KIT

## OPERATING INSTRUCTIONS

## TEST GAS KIT – PART No 99-0010 / 99-0020

---

### 110 LT GAS CAPACITY (1.6LT WATER CAPACITY) – FIXED FLOW REGULATOR

- Ensure regulator control knob is closed (clockwise) before use.
  - Screw onto outlet of the can until it stops and pressure registers on the gauge.
  - The regulator may now be opened by turning the control knob anti-clockwise. Gas will now be delivered at a pre-set flow rate.
  - Always remove regulators after use. All cans will re-seal upon removal, there may be a “hiss” of gas when removal takes place.
  - Attach the test gas applicator attachment to one end of the plastic sample pipe.
  - Connect the other end of the sample pipe to the output nozzle of the regulator.
  - Turn the regulator adjuster anti-clockwise until a flow of gas is obtained from the sample pipe.
  - Screw the test gas applicator onto the sensor.
  - When the gas monitor meter reading has achieved a maximum steady reading adjust the calibration potentiometer for the required meter reading.
  - Remove the test gas applicator and allow the meter to return to zero.
- 
- A decorative graphic consisting of several overlapping, wavy, light gray shapes that resemble stylized waves or smoke, located in the lower right portion of the page.

## GAS MIXTURES

110 LITRE DISPOSABLE CYLINDERS C10 valve  $\frac{5}{8}$ "

PART No	DESCRIPTION
17-0003	Methane 2.5% in air
17-0006	Butane 1% in air
17-0000	Carbon Monoxide 50ppm in air
17-0022	Carbon Dioxide 1% in air
17-0010	Hydrogen Sulphide 10ppm in air
17-0004	Ammonia 25ppm in air
17-0028	Nitric Oxide 25ppm in nitrogen
17-0030	Nitrogen Dioxide 5ppm in air
17-0045	Ethylene Oxide 20ppm in air
17-0016	Nitrogen 100%
17-0018	500ppm R22 in air
Other gases available on request	



Test Gas Kit  
Regulator (fixed flow 0.5 ltr/min) – C10 valve  $\frac{5}{8}$ "  
tube (3M length)  
test gas applicator cap and test gas bung plus carry case  
(does not include test gas)

---

E: [sales@gds-technologies.co.uk](mailto:sales@gds-technologies.co.uk) T: 0113 286 0166 F: 0113 287 8178

GDS TECHNOLOGIES LTD | FUSION POINT | ASH LANE | GARFORTH | LEEDS | LS25 2GA | UK

[www.gds-technologies.co.uk](http://www.gds-technologies.co.uk)